

PROJECT: MAB-X CLINIC

IOM REPORT - OUTPUT-INPUT RISK ANALYSIS

Project Information

Product Type : Drug

TRL : 3

Sprint : 1

Abstract :

mAb-X is a humanized IgG1 monoclonal antibody targeting the interleukin-2 receptor β -chain (CD25) to inhibit T-cell activation and prevent allograft rejection. Early clinical evaluation in renal transplant recipients shows good tolerability and a dose-dependent reduction in activated lymphocytes. The antibody is administered intravenously every two weeks and demonstrates a favorable pharmacokinetic profile with sustained receptor occupancy. Ongoing studies aim to confirm its efficacy and safety in broader transplant populations.

No questions recorded for this sprint.

Start Date : 2025-11-03

End Date : 2026-07-24

Manager

Pr. Thierry Bastogne
julie D

Contributors

François RAGOT
Yasmine Massali
TBgmail

Table of Contents

| | |
|--------------------------------|-----|
| Manufacturing Process Overview | 5 |
| Outputs Identification | N/A |
| FMECA Analysis | 22 |
| Inputs Informations | N/A |
| Cause/Effect Matrix | N/A |

Manufacturing Process Overview

Unit 1: UPSTREAM

Steps:

Step 1.1: Cryovials in Liquid Nitrogen

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|----------------------------------|---------------------------------------------|
| Cell viability after thawing | >90% |
| Identity test | STR profiling or antibody expression marker |
| Mycoplasma testing | <i>No specification</i> |
| Sterility test | <i>No specification</i> |
| Record of freezing/thawing cycle | traceability |

Step 1.2: Cell Revival

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|----------|---------------|
| | |

| | |
|---------------------------------|---------------------------------------|
| Cell count and viability | Trypan Blue or automated cell counter |
| pH and osmolality of the medium | <i>No specification</i> |
| Glucose and lactate levels | <i>No specification</i> |
| Microscopic observation | morphology, contamination check |

Step 1.3: Inoculum Preparation

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|--------------------------|--------------------------------------|
| Viable cell density | target range before transfer |
| Metabolite monitoring | glucose, glutamine, lactate, ammonia |
| Osmolality | <i>No specification</i> |
| Cell morphology | microscopy |
| Absence of contamination | visual or rapid tests |

Step 1.4: Bioreactor Production

Material Attributes:

Feed composition

Basal medium quality

Bioreactor type

Performance Parameters:

pH

Aeration rate & Oxygen flowrate

Feeding mode

Feed rate and schedule

Culture duration

Headspace pressure of the bioreactor

Agitation speed

Temperature

Flowrate of surface air

In-Process Controls (Step)

| IPC Name | Specification |
|----------------------------------------------------|--------------------------------------------------------------------------|
| pH, temperature, DO | real-time probes |
| Viable cell density and viability | Daily sampling - From Output |
| Glucose, lactate, glutamine, ammonia concentration | <i>No specification</i> |
| Product titer | ELISA or HPLC for monitoring yield |
| Agitation speed and aeration rate | Mesures des vraies grandeurs physiques dans le bioréacteur - From Output |
| CO ₂ and O ₂ gas flow rates | <i>No specification</i> |

Step 1.5: Centrifugation

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|----------------------------------------------------------|-------------------------|
| Supernatant turbidity or optical density (clarity check) | clarity check |
| Temperature monitoring | <i>No specification</i> |
| Volume recovery yield | <i>No specification</i> |

Step 1.6: Depth Filtration**Material Attributes:**

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|-----------------------------------------|-------------------------|
| Differential pressure across the filter | <i>No specification</i> |
| Filtration flow rate | <i>No specification</i> |
| Conductivity and pH of filtrate | <i>No specification</i> |

In-Process Controls (IPCs)

| IPC Name | Specification | Step |
|------------------------------|---------------|------------------------------|
| Cell viability after thawing | >90% | Cryovials in Liquid Nitrogen |

| IPC Name | Specification | Step |
|----------------------------------------------------|--------------------------------------------------------------------------|------------------------------|
| Identity test | STR profiling or antibody expression marker | Cryovials in Liquid Nitrogen |
| Mycoplasma testing | <i>No specification</i> | Cryovials in Liquid Nitrogen |
| Sterility test | <i>No specification</i> | Cryovials in Liquid Nitrogen |
| Record of freezing/thawing cycle | traceability | Cryovials in Liquid Nitrogen |
| Cell count and viability | Trypan Blue or automated cell counter | Cell Revival |
| pH and osmolality of the medium | <i>No specification</i> | Cell Revival |
| Glucose and lactate levels | <i>No specification</i> | Cell Revival |
| Microscopic observation | morphology, contamination check | Cell Revival |
| Viable cell density | target range before transfer | Inoculum Preparation |
| Metabolite monitoring | glucose, glutamine, lactate, ammonia | Inoculum Preparation |
| Osmolality | <i>No specification</i> | Inoculum Preparation |
| Cell morphology | microscopy | Inoculum Preparation |
| Absence of contamination | visual or rapid tests | Inoculum Preparation |
| pH, temperature, DO | real-time probes | Bioreactor Production |
| Viable cell density and viability | Daily sampling - From Output | Bioreactor Production |
| Glucose, lactate, glutamine, ammonia concentration | <i>No specification</i> | Bioreactor Production |
| Product titer | ELISA or HPLC for monitoring yield | Bioreactor Production |
| Agitation speed and aeration rate | Mesures des vraies grandeurs physiques dans le bioréacteur - From Output | Bioreactor Production |

| IPC Name | Specification | Step |
|----------------------------------------------------------|------------------------------------------------------|-----------------------|
| CO ₂ and O ₂ gas flow rates | <i>No specification</i> | Bioreactor Production |
| Supernatant turbidity or optical density (clarity check) | clarity check | Centrifugation |
| Temperature monitoring | <i>No specification</i> | Centrifugation |
| Volume recovery yield | <i>No specification</i> | Centrifugation |
| Differential pressure across the filter | <i>No specification</i> | Depth Filtration |
| Filtration flow rate | <i>No specification</i> | Depth Filtration |
| Conductivity and pH of filtrate | <i>No specification</i> | Depth Filtration |
| DO | Dissolved oxygen measured in real time - From Output | <i>No step</i> |

Unit 2: DOWNSTREAM

Steps:

Step 2.1: Capture Chromatography

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|------------------------------------------------|-------------------------|
| UV absorbance (280 nm) to monitor elution peak | <i>No specification</i> |
| Protein concentration in load and eluate | <i>No specification</i> |

| IPC Name | Specification |
|---------------------------------|-------------------------|
| Pressure drop across the column | <i>No specification</i> |

Step 2.2: Viral Inactivation

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|--------------------------------------------------------------------|-------------------------|
| pH measurement and control during inactivation | <i>No specification</i> |
| Product integrity (via HPLC or UV absorbance) after neutralization | <i>No specification</i> |

Step 2.3: Intermediate Chromatography

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|---------------------------------|-------------------------|
| UV absorbance during elution | <i>No specification</i> |
| Conductivity and pH monitoring | <i>No specification</i> |
| Pressure drop across the column | <i>No specification</i> |

Step 2.4: Polishing Chromatography

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|-----------------------------------------|-------------------------|
| UV absorbance (protein peak separation) | <i>No specification</i> |
| Product purity (SDS-PAGE, SEC-HPLC) | <i>No specification</i> |
| Conductivity and pressure | <i>No specification</i> |
| Yield and concentration | <i>No specification</i> |

Step 2.5: Viral Filtration

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|------------------------------------------|-------------------------|
| Differential pressure (?P) across filter | <i>No specification</i> |
| Filtrate turbidity and clarity | <i>No specification</i> |
| Filtrate flow rate | <i>No specification</i> |
| Filter integrity test (before/after run) | <i>No specification</i> |

Step 2.6: Concentration / Diafiltration (TFF)

Material Attributes:*No Material Attributes available***Performance Parameters:***No Performance Parameters available***In-Process Controls (Step)**

| IPC Name | Specification |
|--------------------------------------------------|-------------------------|
| Permeate and retentate flow rates | <i>No specification</i> |
| TMP monitoring | <i>No specification</i> |
| Protein concentration (UV, BCA) | <i>No specification</i> |
| Conductivity and pH (to confirm buffer exchange) | <i>No specification</i> |

Step 2.7: Sterile Filtration (0.22 µm)**Material Attributes:***No Material Attributes available***Performance Parameters:***No Performance Parameters available***In-Process Controls (Step)**

| IPC Name | Specification |
|---------------------------------|-------------------------|
| Filter integrity test | <i>No specification</i> |
| Filtration pressure and time | <i>No specification</i> |
| Protein concentration and yield | <i>No specification</i> |

In-Process Controls (IPCs)

| IPC Name | Specification | Step |
|------------------------------------------------|-------------------------|------------------------|
| UV absorbance (280 nm) to monitor elution peak | <i>No specification</i> | Capture Chromatography |

| IPC Name | Specification | Step |
|--------------------------------------------------------------------|-------------------------|-------------------------------------|
| Protein concentration in load and eluate | <i>No specification</i> | Capture Chromatography |
| Pressure drop across the column | <i>No specification</i> | Capture Chromatography |
| pH measurement and control during inactivation | <i>No specification</i> | Viral Inactivation |
| Product integrity (via HPLC or UV absorbance) after neutralization | <i>No specification</i> | Viral Inactivation |
| UV absorbance during elution | <i>No specification</i> | Intermediate Chromatography |
| Conductivity and pH monitoring | <i>No specification</i> | Intermediate Chromatography |
| Pressure drop across the column | <i>No specification</i> | Intermediate Chromatography |
| UV absorbance (protein peak separation) | <i>No specification</i> | Polishing Chromatography |
| Product purity (SDS-PAGE, SEC-HPLC) | <i>No specification</i> | Polishing Chromatography |
| Conductivity and pressure | <i>No specification</i> | Polishing Chromatography |
| Yield and concentration | <i>No specification</i> | Polishing Chromatography |
| Differential pressure (?P) across filter | <i>No specification</i> | Viral Filtration |
| Filtrate turbidity and clarity | <i>No specification</i> | Viral Filtration |
| Filtrate flow rate | <i>No specification</i> | Viral Filtration |
| Filter integrity test (before/after run) | <i>No specification</i> | Viral Filtration |
| Permeate and retentate flow rates | <i>No specification</i> | Concentration / Diafiltration (TFF) |

| IPC Name | Specification | Step |
|--------------------------------------------------|-------------------------|-------------------------------------|
| TMP monitoring | <i>No specification</i> | Concentration / Diafiltration (TFF) |
| Protein concentration (UV, BCA) | <i>No specification</i> | Concentration / Diafiltration (TFF) |
| Conductivity and pH (to confirm buffer exchange) | <i>No specification</i> | Concentration / Diafiltration (TFF) |
| Filter integrity test | <i>No specification</i> | Sterile Filtration (0.22 µm) |
| Filtration pressure and time | <i>No specification</i> | Sterile Filtration (0.22 µm) |
| Protein concentration and yield | <i>No specification</i> | Sterile Filtration (0.22 µm) |

Unit 3: Fill & Finish

Steps:

Step 3.1: Formulation & Sterile Filtration

Material Attributes:

No Material Attributes available

Performance Parameters:

No Performance Parameters available

In-Process Controls (Step)

| IPC Name | Specification |
|---------------------------------------|-------------------------|
| Filter integrity test | pre- and post-use |
| Protein concentration (UV or HPLC) | <i>No specification</i> |
| Sterility testing (aseptic assurance) | <i>No specification</i> |

Step 3.2: Filling**Material Attributes:***No Material Attributes available***Performance Parameters:***No Performance Parameters available***In-Process Controls (Step)**

| IPC Name | Specification |
|-------------------------------------------------------|--------------------------|
| Sterility and cleanliness of components | <i>No specification</i> |
| In-line weight check or volume control | fill weight verification |
| Container closure integrity testing (CCIT) | <i>No specification</i> |
| Sterility and aseptic process simulation (media fill) | <i>No specification</i> |

Step 3.3: Quality Control & Packaging**Material Attributes:***No Material Attributes available***Performance Parameters:***No Performance Parameters available***In-Process Controls (Step)**

| IPC Name | Specification |
|---------------------------------------|-------------------------|
| Visual inspection of all units | <i>No specification</i> |
| Label verification (OCR/barcode scan) | <i>No specification</i> |
| Stability sampling | <i>No specification</i> |

In-Process Controls (IPCs)

| IPC Name | Specification | Step |
|----------|---------------|------|
| | | |

| | | |
|-------------------------------------------------------|--------------------------|----------------------------------|
| Filter integrity test | pre- and post-use | Formulation & Sterile Filtration |
| Protein concentration (UV or HPLC) | <i>No specification</i> | Formulation & Sterile Filtration |
| Sterility testing (aseptic assurance) | <i>No specification</i> | Formulation & Sterile Filtration |
| Sterility and cleanliness of components | <i>No specification</i> | Filling |
| In-line weight check or volume control | fill weight verification | Filling |
| Container closure integrity testing (CCIT) | <i>No specification</i> | Filling |
| Sterility and aseptic process simulation (media fill) | <i>No specification</i> | Filling |
| Visual inspection of all units | <i>No specification</i> | Quality Control & Packaging |
| Label verification (OCR/barcode scan) | <i>No specification</i> | Quality Control & Packaging |
| Stability sampling | <i>No specification</i> | Quality Control & Packaging |

In-Process Controls (IPCs)

Unit 1 : UPSTREAM

| IPC Name | Specification | Step |
|------------------------------|---------------------------------------------|------------------------------|
| Cell viability after thawing | >90% | Cryovials in Liquid Nitrogen |
| Identity test | STR profiling or antibody expression marker | Cryovials in Liquid Nitrogen |

| IPC Name | Specification | Step |
|----------------------------------------------------|--------------------------------------------------------------------------|------------------------------|
| Mycoplasma testing | <i>No specification</i> | Cryovials in Liquid Nitrogen |
| Sterility test | <i>No specification</i> | Cryovials in Liquid Nitrogen |
| Record of freezing/thawing cycle | traceability | Cryovials in Liquid Nitrogen |
| Cell count and viability | Trypan Blue or automated cell counter | Cell Revival |
| pH and osmolality of the medium | <i>No specification</i> | Cell Revival |
| Glucose and lactate levels | <i>No specification</i> | Cell Revival |
| Microscopic observation | morphology, contamination check | Cell Revival |
| Viable cell density | target range before transfer | Inoculum Preparation |
| Metabolite monitoring | glucose, glutamine, lactate, ammonia | Inoculum Preparation |
| Osmolality | <i>No specification</i> | Inoculum Preparation |
| Cell morphology | microscopy | Inoculum Preparation |
| Absence of contamination | visual or rapid tests | Inoculum Preparation |
| pH, temperature, DO | real-time probes | Bioreactor Production |
| Viable cell density and viability | Daily sampling - From Output | Bioreactor Production |
| Glucose, lactate, glutamine, ammonia concentration | <i>No specification</i> | Bioreactor Production |
| Product titer | ELISA or HPLC for monitoring yield | Bioreactor Production |
| Agitation speed and aeration rate | Mesures des vraies grandeurs physiques dans le bioréacteur - From Output | Bioreactor Production |
| CO ₂ and O ₂ gas flow rates | <i>No specification</i> | Bioreactor Production |

| IPC Name | Specification | Step |
|----------------------------------------------------------|------------------------------------------------------|------------------|
| Supernatant turbidity or optical density (clarity check) | clarity check | Centrifugation |
| Temperature monitoring | <i>No specification</i> | Centrifugation |
| Volume recovery yield | <i>No specification</i> | Centrifugation |
| Differential pressure across the filter | <i>No specification</i> | Depth Filtration |
| Filtration flow rate | <i>No specification</i> | Depth Filtration |
| Conductivity and pH of filtrate | <i>No specification</i> | Depth Filtration |
| DO | Dissolved oxygen measured in real time - From Output | <i>No step</i> |

Unit 2 : DOWNSTREAM

| IPC Name | Specification | Step |
|--------------------------------------------------------------------|-------------------------|-----------------------------|
| UV absorbance (280 nm) to monitor elution peak | <i>No specification</i> | Capture Chromatography |
| Protein concentration in load and eluate | <i>No specification</i> | Capture Chromatography |
| Pressure drop across the column | <i>No specification</i> | Capture Chromatography |
| pH measurement and control during inactivation | <i>No specification</i> | Viral Inactivation |
| Product integrity (via HPLC or UV absorbance) after neutralization | <i>No specification</i> | Viral Inactivation |
| UV absorbance during elution | <i>No specification</i> | Intermediate Chromatography |

| IPC Name | Specification | Step |
|--------------------------------------------------|-------------------------|-------------------------------------|
| Conductivity and pH monitoring | <i>No specification</i> | Intermediate Chromatography |
| Pressure drop across the column | <i>No specification</i> | Intermediate Chromatography |
| UV absorbance (protein peak separation) | <i>No specification</i> | Polishing Chromatography |
| Product purity (SDS-PAGE, SEC-HPLC) | <i>No specification</i> | Polishing Chromatography |
| Conductivity and pressure | <i>No specification</i> | Polishing Chromatography |
| Yield and concentration | <i>No specification</i> | Polishing Chromatography |
| Differential pressure (?P) across filter | <i>No specification</i> | Viral Filtration |
| Filtrate turbidity and clarity | <i>No specification</i> | Viral Filtration |
| Filtrate flow rate | <i>No specification</i> | Viral Filtration |
| Filter integrity test (before/after run) | <i>No specification</i> | Viral Filtration |
| Permeate and retentate flow rates | <i>No specification</i> | Concentration / Diafiltration (TFF) |
| TMP monitoring | <i>No specification</i> | Concentration / Diafiltration (TFF) |
| Protein concentration (UV, BCA) | <i>No specification</i> | Concentration / Diafiltration (TFF) |
| Conductivity and pH (to confirm buffer exchange) | <i>No specification</i> | Concentration / Diafiltration (TFF) |
| Filter integrity test | <i>No specification</i> | Sterile Filtration (0.22 µm) |
| Filtration pressure and time | <i>No specification</i> | Sterile Filtration (0.22 µm) |

| IPC Name | Specification | Step |
|---------------------------------|-------------------------|------------------------------|
| Protein concentration and yield | <i>No specification</i> | Sterile Filtration (0.22 µm) |

Unit 3 : Fill & Finish

| IPC Name | Specification | Step |
|-------------------------------------------------------|--------------------------|----------------------------------|
| Filter integrity test | pre- and post-use | Formulation & Sterile Filtration |
| Protein concentration (UV or HPLC) | <i>No specification</i> | Formulation & Sterile Filtration |
| Sterility testing (aseptic assurance) | <i>No specification</i> | Formulation & Sterile Filtration |
| Sterility and cleanliness of components | <i>No specification</i> | Filling |
| In-line weight check or volume control | fill weight verification | Filling |
| Container closure integrity testing (CCIT) | <i>No specification</i> | Filling |
| Sterility and aseptic process simulation (media fill) | <i>No specification</i> | Filling |
| Visual inspection of all units | <i>No specification</i> | Quality Control & Packaging |
| Label verification (OCR/barcode scan) | <i>No specification</i> | Quality Control & Packaging |
| Stability sampling | <i>No specification</i> | Quality Control & Packaging |

FMECA Analysis :

Scoring Matrix

The scoring matrix is based on the combination of **Uncertainty** and **Severity**.

Higher values indicate a higher impact and risk.

Threshold Low (9) is highlighted in **green**, and Threshold High (25) is highlighted in **red**.

| UNCERTAINTY | SEVERITY | | | |
|-------------|----------|----|----|-----|
| | 1 | 3 | 10 | 25 |
| 1 | 1 | 3 | 10 | 25 |
| 2 | 2 | 6 | 20 | 50 |
| 3 | 3 | 9 | 30 | 75 |
| 4 | 4 | 12 | 40 | 100 |
| 5 | 5 | 15 | 50 | 125 |

Interpretation of the Scoring Matrix

The scoring matrix helps assess the risk associated with different combinations of **Severity** and **Uncertainty**. Below is an explanation of what each combination means:

| Severity | Uncertainty | Interpretation |
|----------------------|-----------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Direct Impact (25) | Limited Data (4) | This combination indicates a critical risk . The severity is high, meaning the issue has a major impact on product quality. Since there is limited data, further investigations are needed to mitigate the risk. |
| Moderate Impact (10) | Clinical Data Available (2) | A moderate risk level. The impact exists, but available clinical data provides confidence in product performance. Regular monitoring is still recommended. |
| Low Impact (3) | Regulatory Data (1) | A low-risk scenario . The process is well-controlled and backed by regulatory data. No immediate concerns, but ongoing verification is advised. |

FMECA Analysis for mAb-X clinic

| OUTPUT NAME | UNIT | STEP | RISK DEFINITION | IMPACTED CATEGORY | SEVERITY | JUSTIFICATIONS | UNCERTAINTY | JUSTIFICATIONS | RPN |
|-----------------------------------------------------------|----------|-----------------------|-----------------|-------------------|-----------------------------|----------------|-------------------------------------------------------------------------|----------------|-----|
| Viable cell density and viability | UPSTREAM | Bioreactor Production | N/A | Quality | Direct impact | N/A | supportive data from clinical studies with this product or similar ones | N/A | 50 |
| Glucose, lactate, glutamine, ammonia concentration | UPSTREAM | Bioreactor Production | N/A | Quality | Moderate or indirect impact | N/A | supportive data from clinical studies with this product or similar ones | N/A | 20 |
| Product titer | UPSTREAM | Bioreactor Production | N/A | Efficacy | Direct impact | N/A | supportive data from clinical studies with this product or similar ones | N/A | 50 |

| OUTPUT NAME | UNIT | STEP | RISK DEFINITION | IMPACTED CATEGORY | SEVERITY | JUSTIFICATIONS | UNCERTAINTY | JUSTIFICATIONS | RPN |
|-------------|----------|-----------------------|-------------------------------------|-------------------|-----------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------|----------------|-----|
| DO | UPSTREAM | Bioreactor Production | Event causing hypoxia and hyperoxia | Quality | Moderate or indirect impact | Hypoxia (Low DO): Reduced cell growth and viability due to insufficient oxygen for cellular respiration. Hyperoxia (High DO): Oxidative stress: Excess oxygen can generate reactive oxygen species (ROS), damaging cells and reducing viability. DO fluctuations can alter the glycosylation pattern of mAbs, affecting effector functions | supportive data from clinical studies with this product or similar ones | N/A | 20 |

Process Analytical Technology (PAT) for mAb-X clinic


| OUTPUT NAME | PROCESS ANALYTICAL TECHNOLOGY | TYPE | TIMING MEASUREMENT | SPECIFICATION (TARGET) | ANALYTICAL PROCEDURES | COMMENT | PROCESS IMPROVEMENTS via PAT |
|----------------------------------------------------|-------------------------------|------|--------------------|------------------------|-----------------------------------------|---------|------------------------------|
| Viable cell density and viability | No PAT available | N/A | N/A | N/A | <i>No Analytical Procedures defined</i> | N/A | N/A |
| Glucose, lactate, glutamine, ammonia concentration | No PAT available | N/A | N/A | N/A | <i>No Analytical Procedures defined</i> | N/A | N/A |
| Product titer | No PAT available | N/A | N/A | N/A | <i>No Analytical Procedures defined</i> | N/A | N/A |
| DO | No PAT available | N/A | N/A | N/A | <i>No Analytical Procedures defined</i> | N/A | N/A |

FMECA Results Interpretation

Criticality Analysis : Pareto Diagram

High Criticality Threshold : 25


Low Criticality Threshold : 9



Criticality Analysis : Histogram

High Criticality Threshold : 25

Low Criticality Threshold : 9




Critical Step Identification for project mAb-X clinic

High Criticality Threshold : **25**

Low Criticality Threshold : **9**

| Project Units | Total Criticality |
|---------------|-------------------|
| UPSTREAM | 140 |
| DOWNSTREAM | 0 |
| Fill & Finish | 0 |

Criticality Analysis for Project



Critical Step Identification per Unit

Unit Name: UPSTREAM


Total Criticality of the Unit: 140

High Criticality Threshold : **25**

Low Criticality Threshold : **9**

| Step Name | Total Criticality |
|------------------------------|-------------------|
| Cryovials in Liquid Nitrogen | 0 |
| Cell Revival | 0 |
| Inoculum Preparation | 0 |
| Bioreactor Production | 140 |
| Centrifugation | 0 |
| Depth Filtration | 0 |

Criticality Analysis for Selected Unit



Unit Name: DOWNSTREAM

Total Criticality of the Unit: 0

High Criticality Threshold : 25

Low Criticality Threshold : 9

| Step Name | Total Criticality |
|-------------------------------------|-------------------|
| Capture Chromatography | 0 |
| Viral Inactivation | 0 |
| Intermediate Chromatography | 0 |
| Polishing Chromatography | 0 |
| Viral Filtration | 0 |
| Concentration / Diafiltration (TFF) | 0 |
| Sterile Filtration (0.22 µm) | 0 |

Criticality Analysis for Selected Unit

Pareto Diagram step/unit

Unit Name: Fill & Finish

Total Criticality of the Unit: 0

High Criticality Threshold : 25

Low Criticality Threshold : 9

| Step Name | Total Criticality |
|----------------------------------|-------------------|
| Formulation & Sterile Filtration | 0 |
| Filling | 0 |

| Step Name | Total Criticality |
|-----------------------------|-------------------|
| Quality Control & Packaging | 0 |

Criticality Analysis for Selected Unit

Pareto Diagram step/unit

Inputs Informations :

Process Parameters

| INPUT NAME | UNIT | STEP | COMMENT |
|--------------------------------------|----------|-----------------------|-------------------------------------------------------------------------|
| pH | UPSTREAM | Bioreactor Production | N/A |
| Aeration rate & Oxygen flowrate | UPSTREAM | Bioreactor Production | DO*=30% NOR=[10;100]% |
| Feeding mode | UPSTREAM | Bioreactor Production | (batch, fed-batch, perfusion). Finally--> fed-batch (bolus) |
| Feed rate and schedule | UPSTREAM | Bioreactor Production | J3, J6 |
| Culture duration | UPSTREAM | Bioreactor Production | Récolte à J10 |
| Headspace pressure of the bioreactor | UPSTREAM | Bioreactor Production | (bar) |
| Agitation speed | UPSTREAM | Bioreactor Production | (rpm) |
| Temperature | UPSTREAM | Bioreactor Production | (°C) |
| Flowrate of surface air | UPSTREAM | Bioreactor Production | To control the amount of foam on the surface. (NL/min), NL=Normo-Litres |

Material Attributes

| INPUT NAME | UNIT | STEP | COMMENT |
|----------------------|----------|-----------------------|--------------------------------------------------------------|
| Feed composition | UPSTREAM | Bioreactor Production | Nutrients, glucose, amino acids, etc. Finally, only glucose. |
| Basal medium quality | UPSTREAM | Bioreactor Production | Composition of the initial medium (standard) |
| Bioreactor type | UPSTREAM | Bioreactor Production | (glass, single-use) |


Cause/Effect Matrix

Cause/Effect Matrix

The Cause/Effect Matrix helps identify relationships between inputs and outputs in the manufacturing process. It highlights critical factors influencing product quality, performance and safety. The table below summarizes key parameters that will be analyzed in the upcoming section.

- Identification of key process inputs
- Correlation between inputs and outputs
- Risk prioritization for process optimization

| Input | Impact on Process | Output |
|---------------------|------------------------------|------------------------------------|
| Material Attributes | Critical for drug stability | Dissolution profile, potency |
| Process Parameters | Influence on bioavailability | Tablet hardness, weight uniformity |
| Material Attributes | Critical for drug stability | Dissolution profile, potency |




Visual representation of process interactions.

Cause-Effect Matrix Analysis for mAb-X clinic

| INPUT NAME | UNIT | STEP | TYPE | Viable cell density and viability (50) | Product titer (50) | RPN | COMMENT | OPERATIONAL RANGE | CONTROL METHOD |
|--------------------------------------|----------|-----------------------|--------------|----------------------------------------|--------------------|------|-------------------------------------------------------------------------|--------------------------------------|-----------------------------------------------|
| Feed composition | UPSTREAM | Bioreactor Production | N/A | | | 531 | Nutrients, glucose, amino acids, etc. Finally, only glucose. | Feed is only composed of glucose | Fixed |
| Basal medium quality | UPSTREAM | Bioreactor Production | N/A | | | 561 | Composition of the initial medium (standard) | Composition of a standard medium | Fixed |
| Bioreactor type | UPSTREAM | Bioreactor Production | N/A | | | 401 | (glass, single-use) | Single-use bioreactor | Fixed |
| pH | UPSTREAM | Bioreactor Production | N/A | | | 711 | N/A | pH* in [6.8 ; 7.2] | Driven by a PID controller |
| Aeration rate & Oxygen flowrate | UPSTREAM | Bioreactor Production | Quantitative | | | 1051 | DO*=30% NOR=[10;100]% | Air: [15-30] NL/min O2: [1-5] L/min | Controlled by Mass Flow Controllers (MFC PID) |
| Feeding mode | UPSTREAM | Bioreactor Production | N/A | | | 381 | (batch, fed-batch, perfusion). Finally--> fed-batch (bolus) | Fed-batch mode | Fixed |
| Feed rate and schedule | UPSTREAM | Bioreactor Production | N/A | | | 651 | J3, J6 | Glucose bolus administered J3 and J6 | Fixed scheme |
| Culture duration | UPSTREAM | Bioreactor Production | Quantitative | | | 651 | Récolte à J10 | Cell harvest at J10 | Fixed |
| Headspace pressure of the bioreactor | UPSTREAM | Bioreactor Production | N/A | | | 81 | (bar) | P*=0.35 bar | Driven by PID controller |
| Agitation speed | UPSTREAM | Bioreactor Production | N/A | | | 576 | (rpm) | S* in [90-110] rpm | driven by a PID controller |
| Temperature | UPSTREAM | Bioreactor Production | N/A | | | 711 | (°C) | T* in [36.5 ; 37.5] °C | Driven by a PID controller |
| Flowrate of surface air | UPSTREAM | Bioreactor Production | N/A | | | 81 | To control the amount of foam on the surface. (NL/min), NL=Normo-Litres | 200 NL/min | fixed value controlled by a MFC |

Cause-Effect Analysis & Pareto Prioritization for mAb-X clinic

RPN Threshold : 600



Input - Output Relation per Input

Input: Feed composition

| INPUTS | OUTPUTS | |
|------------------|-----------------------------------|---------------------------|
| | Viable cell density and viability | Product titer |
| Feed composition | Strong link | Strong link |
| Justifications | No justification provided | No justification provided |

Input: Basal medium quality

| INPUTS | OUTPUTS | |
|----------------------|-----------------------------------|---------------------------|
| | Viable cell density and viability | Product titer |
| Basal medium quality | Perhaps strong or Unknown | Perhaps strong or Unknown |
| Justifications | No justification provided | No justification provided |

Input: Bioreactor type

| INPUTS | OUTPUTS | |
|-----------------|-----------------------------------|---------------------------|
| | Viable cell density and viability | Product titer |
| Bioreactor type | Moderate | Moderate |
| Justifications | No justification provided | No justification provided |

Input: pH

| INPUTS | OUTPUTS | |
|----------------|-----------------------------------|---------------------------|
| | Viable cell density and viability | Product titer |
| pH | Strong link | Strong link |
| Justifications | No justification provided | No justification provided |

Input: Aeration rate & Oxygen flowrate

| INPUTS | OUTPUTS |
|---------------------------------|-------------------------------------------------------------------------|
| Aeration rate & Oxygen flowrate | Viable cell density and viability Strong link |
| Justifications | Product titer Perhaps strong or Unknown No justification provided |

Input: Feeding mode

| INPUTS | OUTPUTS |
|----------------|-------------------------------------------------------------------------|
| Feeding mode | Viable cell density and viability Perhaps strong or Unknown |
| Justifications | Product titer Perhaps strong or Unknown No justification provided |

Input: Feed rate and schedule

| INPUTS | OUTPUTS |
|------------------------|-----------------------------------------------------------|
| Feed rate and schedule | Viable cell density and viability Strong link |
| Justifications | Product titer Strong link No justification provided |

Input: Culture duration

| INPUTS | OUTPUTS |
|------------------|-----------------------------------------------------------|
| Culture duration | Viable cell density and viability Strong link |
| Justifications | Product titer Strong link No justification provided |

Input: Headspace pressure of the bioreactor

| INPUTS | OUTPUTS |
|--------------------------------------|---------------------------------------------------------------|
| Headspace pressure of the bioreactor | Viable cell density and viability No relationship or small |
| Justifications | No justification provided |

Input: Agitation speed

| INPUTS | OUTPUTS |
|-----------------|--------------------------------------------------|
| Agitation speed | Viable cell density and viability Strong link |
| Justifications | No justification provided |

Input: Temperature

| INPUTS | OUTPUTS |
|----------------|--------------------------------------------------|
| Temperature | Viable cell density and viability Strong link |
| Justifications | No justification provided |

Input: Flowrate of surface air

| INPUTS | OUTPUTS |
|-------------------------|---------------------------------------------------------------|
| Flowrate of surface air | Viable cell density and viability No relationship or small |
| Justifications | No justification provided |